



ANALYTICAL REPORT

PREPARED FOR

Attn: Nick Bruzzi
Housatonic Basin Sampling & Testing
80 Run Way
Lee MA 01238

Generated 11/8/2023 8:40 AM

JOB DESCRIPTION

1279000-231010

JOB NUMBER

810-81502-1

Eurofins Eaton Analytical South Bend

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and will meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

Authorization



Generated
11/8/2023 8:40 AM

Authorized for release by
Amanda Scott, Project Manager
Amanda.Scott@et.eurofinsus.com



Per- and Polyfluoroalkyl Substances (PFAS) Report

Page 1 of 2

I. PWS INFORMATION: Please refer to your MassDEP Water Quality Sampling Schedule (WQSS) to help complete this form

PWS ID #:	MA1279000	City / Town:	Southwick
PWS Name:	Southwick Water Department	PWS Class:	COM ☒ NTNC ☐ TNC ☐

MassDEP LOCATION (LOC) ID#	MassDEP Location Name	Sample Information	Date Collected	Collected By
10006	POE Post GT Brook 02G	☐ (M)ultiple ☒ (S)ingle	☐ (R)aw ☒ (F)inished	10/10/2023 Dakota Ketchen
Routine or Special Sample	Original, Resubmitted or Confirmation Report	If Resubmitted Report, list below:		
☒ RS ☐ SS	☒ Original ☐ Resubmitted ☐ Confirmation	(1) Reason for Resubmission	(2) Collection Date of Original Sample	
		☐ Resample ☐ Reanalysis ☐ Report Correction		
SAMPLE COMMENTS — Such as, if a Manifold/Multiple sample, list the source(s) that were on-line during sample collection or if this is a field reagent blank.				

II. ANALYTICAL LABORATORY INFORMATION:

Primary Lab MA Cert. #:	M-IN035	Primary Lab Name:	Eurofins Eaton Analytical South Bend	Subcontracted? (Y/N)	N
Analysis Lab MA Cert. #:	M-IN035	Analysis Lab Name:	Eurofins Eaton Analytical South Bend		
If Analysis Lab is not certified by MassDEP or U.S. EPA, list certification authority:					

Lab Method	Date Extracted	Date Analyzed	Dilution Factor	Lab Sample ID#	
537.1	10/19/2023	10/21/2023	1	Primary Lab:	810-81502-1
				Subcontracted Lab:	810-81502-1

CAS#	REGULATED PFAS CONTAMINANTS	Result ¹ ng/L	Result ² Qualifier	MCL* ng/L	MDL ng/L	MRL ng/L
1763-23-1	Perfluorooctane Sulfonic Acid (PFOS)	0.98	J	-	0.49	1.9
335-67-1	Perfluorooctanoic Acid (PFOA)	1.8	J	-	0.47	1.9
355-46-4	Perfluorohexane Sulfonic Acid (PFHxS)	1.1	J	-	0.41	1.9
375-95-1	Perfluorononanoic Acid (PFNA)	ND		-	0.45	1.9
375-85-9	Perfluoroheptanoic Acid (PFHpA)	0.60	J	-	0.49	1.9
335-76-2	Perfluorodecanoic acid (PFDA)	ND		-	0.56	1.9
PFAS6 (sum of PFOS, PFOA, PFHxS, PFNA, PFHpA and PFDA; only include Results at or above the MRL; do not include estimated Results as described by a Result Qualifier in the next column)		ND	-	20	-	-
UNREGULATED PFAS CONTAMINANTS						
375-73-5	Perfluorobutane sulfonic acid (PFBS)	0.72	J	-	0.66	1.9
307-55-1	Perfluorododecanoic acid (PFDoA)	ND		-	0.59	1.9
307-24-4	Perfluorohexanoic acid (PFHxA)	1.4	J	-	0.59	1.9
376-06-7	Perfluorotetradecanoic acid (PFTA)	ND		-	0.61	1.9
72629-94-8	Perfluorotridecanoic acid (PFTrDA)	ND		-	0.56	1.9
2058-94-8	Perfluoroundecanoic acid (PFUnA)	ND		-	0.59	1.9
2991-50-6	N-ethyl perfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		-	0.48	1.9
2355-31-9	N-methyl perfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		-	0.58	1.9
763051-92-9	11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	ND		-	0.60	1.9
756426-58-1	9-chlorohexadecafluoro-3-oxanone-1-sulfonic acid (9Cl-PF3ONS)	ND		-	0.60	1.9
919005-14-4	4,8-dioxo-3H-perfluorononanoic acid (ADONA)	ND		-	0.46	1.9
13252-13-6	Hexafluoropropylene oxide dimer acid (HFPO-DA)	ND		-	0.58	1.9

¹ A field reagent blank (FRB) must be analyzed and reported on a separate PFAS form if any PFAS are detected above the MRL.² All qualifiers must be described under Lab Analysis Comments on page 2.



Per- and Polyfluoroalkyl Substances (PFAS) Report

Page 2 of 2

PWS ID #:

MA1279000

Lab Sample ID #:

Primary Lab: 810-81502-1

Subcontracted Lab: 810-81502-1

CAS#	UNREGULATED PFAS CONTAMINANTS	Result ¹ ng/L	Result ² Qualifier	MCL* ng/L	MDL ng/L	MRL ng/L
				-		
				-		
				-		
				-		
				-		
				-		
				-		
				-		
				-		
				-		
				-		
				-		
				-		
				-		
				-		

Surrogate Name	% Recovery (70 — 130%)	Alternate Surrogate (must document reason for change)
13C2-PFHxA	86	
13C2-PFDA	88	
d5-NEtFOSAA	89	
13C3-HFPO-DA	88	

Note: 13C3-HFDO-DA is not required for EPA Method 537 v1.1

In addition to the SUR above you must attach the results of the ongoing QC results as specified by the method for the sample's extraction batch.

☒ Laboratory analytical report with QC attached (check one item below).☒ All associated QC criteria reported within control limits including Lab Reagent/Method Blank (LRB), Field Reagent Blank (FRB), Surrogate Standards (SUR), Laboratory Fortified Blank (LFB), Matrix Spike/Duplicate (LFSM/LFSMD or FD) and RPD.☐ All associated sample and/or QC batch criteria not met. See Lab Analysis Comments below and narrative in attached report.

Lab Analysis Comments: (include sample/method parameters outside of or affecting QC controls/limits and result qualifiers)

Result Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
Other Analysis Comments:	

* MCL or proposed MCL

I certify under penalties of law that I am the person authorized to fill out this form and the information contained herein is true, accurate and complete to the best extent of my knowledge.

Authorized Signature:

Amanda Scott, Project Manager

Date: 11/08/2023

If not submitting these results electronically, mail TWO copies of this report to your MassDEP Regional Office no later than 10 days after the end of the month in which you received this report or no later than 10 days after the end of the reporting period, whichever is sooner. Note that during the Massachusetts COVID-19 state of emergency, in addition to submitting by mail, reports may be emailed to program.director-dwp@mass.gov.

MassDEP REVIEW STATUS (Initial and Date)	Review Comments	☐ WQTS Data Entered
☐ Accepted _____ ☐ Disapproved _____		

ANALYTICAL REPORT

PREPARED FOR

Attn: Nick Bruzzi
Housatonic Basin Sampling & Testing
80 Run Way
Lee, Massachusetts 01238

Generated 11/6/2023 7:43:08 PM

JOB DESCRIPTION

1279000-231010

JOB NUMBER

810-81502-1

Eurofins Eaton Analytical South Bend

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and will meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

Authorization



Generated
11/6/2023 7:43:08 PM

Authorized for release by
Amanda Scott, Project Manager
Amanda.Scott@et.eurofinsus.com
(574)233-4777

Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Surrogate Summary	9
QC Sample Results	10
QC Association Summary	15
Lab Chronicle	16
Certification Summary	17
Method Summary	18
Sample Summary	19
Chain of Custody	20
Receipt Checklists	21



Definitions/Glossary

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
α	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Job ID: 810-81502-1

Laboratory: Eurofins Eaton Analytical South Bend

Narrative

Job Narrative 810-81502-1

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers are applied to indicate exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method. Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Receipt

The samples were received on 10/13/2023 9:00 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 0.8°C

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Client Sample ID: POE Post GT Brook 02G

Lab Sample ID: 810-81502-1

PWSID Number: 1279000

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorooctanesulfonic acid (PFOS)	0.98	J	1.9	0.49	ng/L	1		537.1	Total/NA
Perfluorohexanoic acid (PFHxA)	1.4	J	1.9	0.59	ng/L	1		537.1	Total/NA
Perfluorooctanoic acid (PFOA)	1.8	J	1.9	0.47	ng/L	1		537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	1.1	J	1.9	0.41	ng/L	1		537.1	Total/NA
Perfluorobutanesulfonic acid (PFBS)	0.72	J	1.9	0.66	ng/L	1		537.1	Total/NA
Perfluoroheptanoic acid (PFHpA)	0.60	J	1.9	0.49	ng/L	1		537.1	Total/NA

Client Sample ID: POE Post GT Brook 02G FTB

Lab Sample ID: 810-81502-2

PWSID Number: 1279000

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Client Sample ID: POE Post GT Brook 02G

Lab Sample ID: 810-81502-1

Date Collected: 10/10/23 09:35

Matrix: Drinking Water

Date Received: 10/13/23 09:00

PWSID Number: 1279000

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	0.98	J	1.9	0.49	ng/L		10/19/23 15:39	10/21/23 03:28	1
Perfluoroundecanoic acid (PFUnA)	ND		1.9	0.59	ng/L		10/19/23 15:39	10/21/23 03:28	1
Perfluorohexanoic acid (PFHxA)	1.4	J	1.9	0.59	ng/L		10/19/23 15:39	10/21/23 03:28	1
Perfluorododecanoic acid (PFDoA)	ND		1.9	0.59	ng/L		10/19/23 15:39	10/21/23 03:28	1
Perfluorooctanoic acid (PFOA)	1.8	J	1.9	0.47	ng/L		10/19/23 15:39	10/21/23 03:28	1
Perfluorodecanoic acid (PFDA)	ND		1.9	0.56	ng/L		10/19/23 15:39	10/21/23 03:28	1
Perfluorohexanesulfonic acid (PFHxS)	1.1	J	1.9	0.41	ng/L		10/19/23 15:39	10/21/23 03:28	1
Perfluorobutanesulfonic acid (PFBS)	0.72	J	1.9	0.66	ng/L		10/19/23 15:39	10/21/23 03:28	1
Perfluoroheptanoic acid (PFHpA)	0.60	J	1.9	0.49	ng/L		10/19/23 15:39	10/21/23 03:28	1
Perfluorononanoic acid (PFNA)	ND		1.9	0.45	ng/L		10/19/23 15:39	10/21/23 03:28	1
Perfluorotetradecanoic acid (PFTeDA)	ND		1.9	0.61	ng/L		10/19/23 15:39	10/21/23 03:28	1
Perfluorotridecanoic acid (PFTrDA)	ND		1.9	0.56	ng/L		10/19/23 15:39	10/21/23 03:28	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		1.9	0.58	ng/L		10/19/23 15:39	10/21/23 03:28	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		1.9	0.48	ng/L		10/19/23 15:39	10/21/23 03:28	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		1.9	0.58	ng/L		10/19/23 15:39	10/21/23 03:28	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	ND		1.9	0.60	ng/L		10/19/23 15:39	10/21/23 03:28	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	ND		1.9	0.60	ng/L		10/19/23 15:39	10/21/23 03:28	1
4,8-Dioxo-3H-perfluorononanoic acid (ADONA)	ND		1.9	0.46	ng/L		10/19/23 15:39	10/21/23 03:28	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	86		70 - 130				10/19/23 15:39	10/21/23 03:28	1
13C2 PFDA	88		70 - 130				10/19/23 15:39	10/21/23 03:28	1
13C3 HFPO-DA	88		70 - 130				10/19/23 15:39	10/21/23 03:28	1
d5-NEtFOSAA	89		70 - 130				10/19/23 15:39	10/21/23 03:28	1

Method: EPA PFAS6 - PFAS6

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
PFAS Total	ND		2.00	0.500	ng/L			10/21/23 03:28	1

Client Sample ID: POE Post GT Brook 02G FTB

Lab Sample ID: 810-81502-2

Date Collected: 10/10/23 09:35

Matrix: Drinking Water

Date Received: 10/13/23 09:00

PWSID Number: 1279000

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	ND		1.9	0.51	ng/L		10/24/23 07:43	10/25/23 07:20	1
Perfluoroundecanoic acid (PFUnA)	ND		1.9	0.61	ng/L		10/24/23 07:43	10/25/23 07:20	1
Perfluorohexanoic acid (PFHxA)	ND		1.9	0.61	ng/L		10/24/23 07:43	10/25/23 07:20	1
Perfluorododecanoic acid (PFDoA)	ND		1.9	0.61	ng/L		10/24/23 07:43	10/25/23 07:20	1
Perfluorooctanoic acid (PFOA)	ND		1.9	0.49	ng/L		10/24/23 07:43	10/25/23 07:20	1
Perfluorodecanoic acid (PFDA)	ND		1.9	0.58	ng/L		10/24/23 07:43	10/25/23 07:20	1
Perfluorohexanesulfonic acid (PFHxS)	ND		1.9	0.43	ng/L		10/24/23 07:43	10/25/23 07:20	1
Perfluorobutanesulfonic acid (PFBS)	ND		1.9	0.69	ng/L		10/24/23 07:43	10/25/23 07:20	1
Perfluoroheptanoic acid (PFHpA)	ND		1.9	0.50	ng/L		10/24/23 07:43	10/25/23 07:20	1

Eurofins Eaton Analytical South Bend

Client Sample Results

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Client Sample ID: POE Post GT Brook 02G FTB

Lab Sample ID: 810-81502-2

Date Collected: 10/10/23 09:35

Matrix: Drinking Water

Date Received: 10/13/23 09:00

PWSID Number: 1279000

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorononanoic acid (PFNA)	ND		1.9	0.47	ng/L		10/24/23 07:43	10/25/23 07:20	1
Perfluorotetradecanoic acid (PFTeDA)	ND		1.9	0.63	ng/L		10/24/23 07:43	10/25/23 07:20	1
Perfluorotridecanoic acid (PFTrDA)	ND		1.9	0.58	ng/L		10/24/23 07:43	10/25/23 07:20	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		1.9	0.60	ng/L		10/24/23 07:43	10/25/23 07:20	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		1.9	0.50	ng/L		10/24/23 07:43	10/25/23 07:20	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		1.9	0.60	ng/L		10/24/23 07:43	10/25/23 07:20	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	ND		1.9	0.62	ng/L		10/24/23 07:43	10/25/23 07:20	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	ND		1.9	0.62	ng/L		10/24/23 07:43	10/25/23 07:20	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	ND		1.9	0.48	ng/L		10/24/23 07:43	10/25/23 07:20	1
Surrogate	%Recovery	Qualifier	Limits				Prepared	Analyzed	Dil Fac
13C2 PFHxA	102		70 - 130				10/24/23 07:43	10/25/23 07:20	1
13C2 PFDA	104		70 - 130				10/24/23 07:43	10/25/23 07:20	1
13C3 HFPO-DA	102		70 - 130				10/24/23 07:43	10/25/23 07:20	1
d5-NEtFOSAA	106		70 - 130				10/24/23 07:43	10/25/23 07:20	1

Method: EPA PFAS6 - PFAS6

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
PFAS Total	ND		2.00	0.500	ng/L			10/25/23 07:20	1

Surrogate Summary

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		PFHxA (70-130)	PFDA (70-130)	HFPODA (70-130)	d5NEFOS (70-130)
810-81502-1	POE Post GT Brook 02G	86	88	88	89
810-81502-1 DU	POE Post GT Brook 02G	85	90	89	84
810-81502-2	POE Post GT Brook 02G FTB	102	104	102	106
LCS 810-77409/3-A	Lab Control Sample	93	92	91	92
LLCS 810-77409/2-A	Lab Control Sample	86	86	83	92
LLCS 810-77704/2-A	Lab Control Sample	93	98	90	101
MBL 810-77409/1-A	Method Blank	84	87	85	89
MBL 810-77704/1-A	Method Blank	97	100	97	101

Surrogate Legend

PFHxA = 13C2 PFHxA

PFDA = 13C2 PFDA

HFPODA = 13C3 HFPO-DA

d5NEFOS = d5-NEtFOSAA

QC Sample Results

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MBL 810-77409/1-A

Matrix: Drinking Water

Analysis Batch: 77549

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 77409

Analyte	MBL Result	MBL Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	ND		2.0	0.53	ng/L		10/19/23 15:39	10/21/23 02:35	1
Perfluoroundecanoic acid (PFUnA)	ND		2.0	0.63	ng/L		10/19/23 15:39	10/21/23 02:35	1
Perfluorohexanoic acid (PFHxA)	ND		2.0	0.63	ng/L		10/19/23 15:39	10/21/23 02:35	1
Perfluorododecanoic acid (PFDoA)	ND		2.0	0.63	ng/L		10/19/23 15:39	10/21/23 02:35	1
Perfluorooctanoic acid (PFOA)	ND		2.0	0.50	ng/L		10/19/23 15:39	10/21/23 02:35	1
Perfluorodecanoic acid (PFDA)	ND		2.0	0.60	ng/L		10/19/23 15:39	10/21/23 02:35	1
Perfluorohexanesulfonic acid (PFHxS)	ND		2.0	0.44	ng/L		10/19/23 15:39	10/21/23 02:35	1
Perfluorobutanesulfonic acid (PFBS)	ND		2.0	0.71	ng/L		10/19/23 15:39	10/21/23 02:35	1
Perfluoroheptanoic acid (PFHpA)	ND		2.0	0.52	ng/L		10/19/23 15:39	10/21/23 02:35	1
Perfluorononanoic acid (PFNA)	ND		2.0	0.48	ng/L		10/19/23 15:39	10/21/23 02:35	1
Perfluorotetradecanoic acid (PFTeDA)	ND		2.0	0.65	ng/L		10/19/23 15:39	10/21/23 02:35	1
Perfluorotridecanoic acid (PFTriDA)	ND		2.0	0.60	ng/L		10/19/23 15:39	10/21/23 02:35	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		2.0	0.62	ng/L		10/19/23 15:39	10/21/23 02:35	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		2.0	0.51	ng/L		10/19/23 15:39	10/21/23 02:35	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		2.0	0.62	ng/L		10/19/23 15:39	10/21/23 02:35	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	ND		2.0	0.64	ng/L		10/19/23 15:39	10/21/23 02:35	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	ND		2.0	0.64	ng/L		10/19/23 15:39	10/21/23 02:35	1
4,8-Dioxo-3H-perfluorononanoic acid (ADONA)	ND		2.0	0.49	ng/L		10/19/23 15:39	10/21/23 02:35	1

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFHxA	84		70 - 130	10/19/23 15:39	10/21/23 02:35	1
13C2 PFDA	87		70 - 130	10/19/23 15:39	10/21/23 02:35	1
13C3 HFPO-DA	85		70 - 130	10/19/23 15:39	10/21/23 02:35	1
d5-NEtFOSAA	89		70 - 130	10/19/23 15:39	10/21/23 02:35	1

Lab Sample ID: LCS 810-77409/3-A

Matrix: Drinking Water

Analysis Batch: 77549

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 77409

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluorooctanesulfonic acid (PFOS)	100	81.8		ng/L		82	70 - 130
Perfluoroundecanoic acid (PFUnA)	100	80.6		ng/L		81	70 - 130
Perfluorohexanoic acid (PFHxA)	100	84.4		ng/L		84	70 - 130
Perfluorododecanoic acid (PFDoA)	100	83.5		ng/L		84	70 - 130
Perfluorooctanoic acid (PFOA)	100	81.5		ng/L		82	70 - 130
Perfluorodecanoic acid (PFDA)	100	84.8		ng/L		85	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	100	84.2		ng/L		84	70 - 130
Perfluorobutanesulfonic acid (PFBS)	100	78.2		ng/L		78	70 - 130
Perfluoroheptanoic acid (PFHpA)	100	89.4		ng/L		89	70 - 130
Perfluorononanoic acid (PFNA)	100	85.4		ng/L		85	70 - 130

Eurofins Eaton Analytical South Bend

QC Sample Results

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: LCS 810-77409/3-A

Matrix: Drinking Water

Analysis Batch: 77549

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 77409

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluorotetradecanoic acid (PFTeDA)	100	72.2		ng/L		72	70 - 130
Perfluorotridecanoic acid (PFTrDA)	100	81.5		ng/L		81	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	100	78.6		ng/L		79	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	100	82.5		ng/L		82	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	100	82.5		ng/L		83	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	100	81.1		ng/L		81	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	100	79.4		ng/L		79	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	100	88.5		ng/L		89	70 - 130
Surrogate	%Recovery	LCS Qualifier	LCS	Limits			
13C2 PFHxA	93			70 - 130			
13C2 PFDA	92			70 - 130			
13C3 HFPO-DA	91			70 - 130			
d5-NEtFOSAA	92			70 - 130			

Lab Sample ID: LLCS 810-77409/2-A

Matrix: Drinking Water

Analysis Batch: 77549

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 77409

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluorooctanesulfonic acid (PFOS)	2.00	1.51	J	ng/L		75	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	1.52	J	ng/L		76	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	1.68	J	ng/L		84	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	1.64	J	ng/L		82	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	1.58	J	ng/L		79	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	1.64	J	ng/L		82	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	1.59	J	ng/L		80	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	1.42	J	ng/L		71	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	1.62	J	ng/L		81	50 - 150
Perfluorononanoic acid (PFNA)	2.00	1.64	J	ng/L		82	50 - 150
Perfluorotetradecanoic acid (PFTeDA)	2.00	1.35	J	ng/L		67	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.00	1.52	J	ng/L		76	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.00	1.47	J	ng/L		73	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.00	1.62	J	ng/L		81	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	2.00	1.49	J	ng/L		74	50 - 150

Eurofins Eaton Analytical South Bend

QC Sample Results

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: LLCS 810-77409/2-A

Matrix: Drinking Water

Analysis Batch: 77549

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 77409

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits
9-Chlorohexadecafluoro-3-oxan onane-1-sulfonic acid	2.00	1.52	J	ng/L		76	50 - 150
11-Chloroeicosafluoro-3-oxaund ecane-1-sulfonic acid	2.00	1.37	J	ng/L		69	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	1.64	J	ng/L		82	50 - 150

Surrogate	LLCS %Recovery	LLCS Qualifier	Limits
13C2 PFHxA	86		70 - 130
13C2 PFDA	86		70 - 130
13C3 HFPO-DA	83		70 - 130
d5-NEtFOSAA	92		70 - 130

Lab Sample ID: 810-81502-1 DU

Matrix: Drinking Water

Analysis Batch: 77549

Client Sample ID: POE Post GT Brook 02G

Prep Type: Total/NA

Prep Batch: 77409

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	Limit
Perfluorooctanesulfonic acid (PFOS)	0.98	J	1.02	J	ng/L		4	30
Perfluoroundecanoic acid (PFUnA)	ND		ND		ng/L		NC	30
Perfluorohexanoic acid (PFHxA)	1.4	J	1.42	J	ng/L		1	30
Perfluorododecanoic acid (PFDoA)	ND		ND		ng/L		NC	30
Perfluorooctanoic acid (PFOA)	1.8	J	1.84	J	ng/L		0.9	30
Perfluorodecanoic acid (PFDA)	ND		ND		ng/L		NC	30
Perfluorohexanesulfonic acid (PFHxS)	1.1	J	1.22	J	ng/L		6	30
Perfluorobutanesulfonic acid (PFBS)	0.72	J	0.679	J	ng/L		6	30
Perfluoroheptanoic acid (PFHpA)	0.60	J	0.607	J	ng/L		1	30
Perfluorononanoic acid (PFNA)	ND		ND		ng/L		NC	30
Perfluorotetradecanoic acid (PFTeDA)	ND		ND		ng/L		NC	30
Perfluorotridecanoic acid (PFTrDA)	ND		ND		ng/L		NC	30
N-methylperfluorooctanesulfona midoacetic acid (NMeFOSAA)	ND		ND		ng/L		NC	30
N-ethylperfluorooctanesulfonami doacetic acid (NEtFOSAA)	ND		ND		ng/L		NC	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		ND		ng/L		NC	30
9-Chlorohexadecafluoro-3-oxan onane-1-sulfonic acid	ND		ND		ng/L		NC	30
11-Chloroeicosafluoro-3-oxaund ecane-1-sulfonic acid	ND		ND		ng/L		NC	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	ND		ND		ng/L		NC	30

Surrogate	DU %Recovery	DU Qualifier	Limits
13C2 PFHxA	85		70 - 130

Eurofins Eaton Analytical South Bend

QC Sample Results

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: 810-81502-1 DU

Matrix: Drinking Water

Analysis Batch: 77549

Client Sample ID: POE Post GT Brook 02G

Prep Type: Total/NA

Prep Batch: 77409

Surrogate	%Recovery	DU DU Qualifier	Limits
13C2 PFDA	90		70 - 130
13C3 HFPO-DA	89		70 - 130
d5-NEtFOSAA	84		70 - 130

Lab Sample ID: MBL 810-77704/1-A

Matrix: Drinking Water

Analysis Batch: 77772

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 77704

Analyte	MBL Result	MBL Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	ND		2.0	0.53	ng/L		10/24/23 07:43	10/24/23 23:02	1
Perfluoroundecanoic acid (PFUnA)	ND		2.0	0.63	ng/L		10/24/23 07:43	10/24/23 23:02	1
Perfluorohexanoic acid (PFHxA)	ND		2.0	0.63	ng/L		10/24/23 07:43	10/24/23 23:02	1
Perfluorododecanoic acid (PFDoA)	ND		2.0	0.63	ng/L		10/24/23 07:43	10/24/23 23:02	1
Perfluorooctanoic acid (PFOA)	ND		2.0	0.50	ng/L		10/24/23 07:43	10/24/23 23:02	1
Perfluorodecanoic acid (PFDA)	ND		2.0	0.60	ng/L		10/24/23 07:43	10/24/23 23:02	1
Perfluorohexanesulfonic acid (PFHxS)	ND		2.0	0.44	ng/L		10/24/23 07:43	10/24/23 23:02	1
Perfluorobutanesulfonic acid (PFBS)	ND		2.0	0.71	ng/L		10/24/23 07:43	10/24/23 23:02	1
Perfluoroheptanoic acid (PFHpA)	ND		2.0	0.52	ng/L		10/24/23 07:43	10/24/23 23:02	1
Perfluorononanoic acid (PFNA)	ND		2.0	0.48	ng/L		10/24/23 07:43	10/24/23 23:02	1
Perfluorotetradecanoic acid (PFTeDA)	ND		2.0	0.65	ng/L		10/24/23 07:43	10/24/23 23:02	1
Perfluorotridecanoic acid (PFTTrDA)	ND		2.0	0.60	ng/L		10/24/23 07:43	10/24/23 23:02	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		2.0	0.62	ng/L		10/24/23 07:43	10/24/23 23:02	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		2.0	0.51	ng/L		10/24/23 07:43	10/24/23 23:02	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		2.0	0.62	ng/L		10/24/23 07:43	10/24/23 23:02	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	ND		2.0	0.64	ng/L		10/24/23 07:43	10/24/23 23:02	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	ND		2.0	0.64	ng/L		10/24/23 07:43	10/24/23 23:02	1
4,8-Dioxo-3H-perfluorononanoic acid (ADONA)	ND		2.0	0.49	ng/L		10/24/23 07:43	10/24/23 23:02	1

Surrogate	%Recovery	MBL MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFHxA	97		70 - 130	10/24/23 07:43	10/24/23 23:02	1
13C2 PFDA	100		70 - 130	10/24/23 07:43	10/24/23 23:02	1
13C3 HFPO-DA	97		70 - 130	10/24/23 07:43	10/24/23 23:02	1
d5-NEtFOSAA	101		70 - 130	10/24/23 07:43	10/24/23 23:02	1

Lab Sample ID: LLCS 810-77704/2-A

Matrix: Drinking Water

Analysis Batch: 77772

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 77704

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluorooctanesulfonic acid (PFOS)	2.00	1.91	J	ng/L		95	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	1.91	J	ng/L		95	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	1.94	J	ng/L		97	50 - 150

Eurofins Eaton Analytical South Bend

QC Sample Results

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: LLCS 810-77704/2-A

Matrix: Drinking Water

Analysis Batch: 77772

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 77704

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluorododecanoic acid (PFDoA)	2.00	1.89	J	ng/L		95	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	1.88	J	ng/L		94	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	1.95	J	ng/L		97	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.01		ng/L		100	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	1.79	J	ng/L		89	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.07		ng/L		104	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.07		ng/L		103	50 - 150
Perfluorotetradecanoic acid (PFTeDA)	2.00	1.73	J	ng/L		87	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.00	1.84	J	ng/L		92	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.00	1.80	J	ng/L		90	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.00	1.98	J	ng/L		99	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	2.00	1.81	J	ng/L		91	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	2.00	1.93	J	ng/L		97	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	2.00	1.84	J	ng/L		92	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	1.95	J	ng/L		98	50 - 150

Surrogate	LLCS %Recovery	LLCS Qualifier	Limits
13C2 PFHxA	93		70 - 130
13C2 PFDA	98		70 - 130
13C3 HFPO-DA	90		70 - 130
d5-NEtFOSAA	101		70 - 130

QC Association Summary

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

LCMS

Prep Batch: 77409

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-81502-1	POE Post GT Brook 02G	Total/NA	Drinking Water	537.1 DW	
MBL 810-77409/1-A	Method Blank	Total/NA	Drinking Water	537.1 DW	
LCS 810-77409/3-A	Lab Control Sample	Total/NA	Drinking Water	537.1 DW	
LLCS 810-77409/2-A	Lab Control Sample	Total/NA	Drinking Water	537.1 DW	
810-81502-1 DU	POE Post GT Brook 02G	Total/NA	Drinking Water	537.1 DW	

Analysis Batch: 77549

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-81502-1	POE Post GT Brook 02G	Total/NA	Drinking Water	537.1	77409
MBL 810-77409/1-A	Method Blank	Total/NA	Drinking Water	537.1	77409
LCS 810-77409/3-A	Lab Control Sample	Total/NA	Drinking Water	537.1	77409
LLCS 810-77409/2-A	Lab Control Sample	Total/NA	Drinking Water	537.1	77409
810-81502-1 DU	POE Post GT Brook 02G	Total/NA	Drinking Water	537.1	77409

Prep Batch: 77704

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-81502-2	POE Post GT Brook 02G FTB	Total/NA	Drinking Water	537.1 DW	
MBL 810-77704/1-A	Method Blank	Total/NA	Drinking Water	537.1 DW	
LLCS 810-77704/2-A	Lab Control Sample	Total/NA	Drinking Water	537.1 DW	

Analysis Batch: 77772

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-81502-2	POE Post GT Brook 02G FTB	Total/NA	Drinking Water	537.1	77704
MBL 810-77704/1-A	Method Blank	Total/NA	Drinking Water	537.1	77704
LLCS 810-77704/2-A	Lab Control Sample	Total/NA	Drinking Water	537.1	77704

Analysis Batch: 78820

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-81502-1	POE Post GT Brook 02G	Total/NA	Drinking Water	PFAS6	
810-81502-2	POE Post GT Brook 02G FTB	Total/NA	Drinking Water	PFAS6	

Lab Chronicle

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Client Sample ID: POE Post GT Brook 02G

Lab Sample ID: 810-81502-1

Date Collected: 10/10/23 09:35

Matrix: Drinking Water

Date Received: 10/13/23 09:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	537.1 DW			77409	EH	EA SB	10/19/23 15:39
Total/NA	Analysis	537.1		1	77549	UKDT	EA SB	10/21/23 03:28
Total/NA	Analysis	PFAS6		1	78820	JT	EA SB	10/21/23 03:28

Client Sample ID: POE Post GT Brook 02G FTB

Lab Sample ID: 810-81502-2

Date Collected: 10/10/23 09:35

Matrix: Drinking Water

Date Received: 10/13/23 09:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	537.1 DW			77704	MP	EA SB	10/24/23 07:43
Total/NA	Analysis	537.1		1	77772	UKDT	EA SB	10/25/23 07:20
Total/NA	Analysis	PFAS6		1	78820	JT	EA SB	10/25/23 07:20

Laboratory References:

EA SB = Eurofins Eaton Analytical South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777

Accreditation/Certification Summary

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Laboratory: Eurofins Eaton Analytical South Bend

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
Massachusetts	State	M-IN035	06-30-24

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
PFAS6		Drinking Water	PFAS Total

Method Summary

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010

Job ID: 810-81502-1

Method	Method Description	Protocol	Laboratory
537.1	Perfluorinated Alkyl Acids (LC/MS)	EPA	EA SB
PFAS6	PFAS6	EPA	EA SB
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA SB

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA SB = Eurofins Eaton Analytical South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777

Sample Summary

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000-231010




Job ID: 810-81502-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
810-81502-1	POE Post GT Brook 02G	Drinking Water	10/10/23 09:35	10/13/23 09:00	1279000
810-81502-2	POE Post GT Brook 02G FTB	Drinking Water	10/10/23 09:35	10/13/23 09:00	1279000

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15

SAMPLE COLLECTION RECORD | CHAIN OF CUSTODY | EUROFINIS EATON ANALYTICAL

PWS NAME: SOUTHWICK WATER DEPARTMENT				80 RUN WAY	
PWS ID: 1279000	PWS TOWN: Southwick			LEE, MA 01238	
PWS CLASS: COM				(413)248-4622	
		HBST P.O. # 1279000-231010	# of WC:	###	

SAMPLE INFORMATION				FIELD RECORDED			MICRO BIOLOGY		CHEMICAL ANALYSIS																
ID	SAMPLE TYPE	BACTERIA DEP ID	Chem Sample ID	LOCATION DESCRIPTION	DATE/TIME	SAMPLER	Field Temp C*	Field pH	Field Turbidity (NTU)	Chi2 Res (Free)	BACTERIA HPC	BACTERIA 9223 P/A	BACTERIA 9223 QT	PFAS 537.1	PFAS Blank										
1100061				POE POST GT BROOK 02G -	10/10/23 9:35 AM	Dakota Keichien								X	X										
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					
				 810-81502 Chain of Custody																					

1a. Submit via EDEP unless designated Private or otherwise noted. Email report to: Admin@HousatonicBasin.com. Lab testing shall be in compliance with all State and Federal Drinking Water and applicable regulations.

Login Sample Receipt Checklist

Client: Housatonic Basin Sampling & Testing

Job Number: 810-81502-1

Login Number: 81502

List Source: Eurofins Eaton Analytical South Bend

List Number: 1

Creator: DePriest, Kellie

Question	Answer	Comment
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	