eurofins | Environment Testing

ANALYTICAL REPORT

Eurofins Eaton South Bend
110 S Hill Street
South Bend, IN 46617
Tel: (574)233-4777

Laboratory Job ID: 810-41842-1
Client Project/Site: 1279000

For:
Housatonic Basin Sampling & Testing
80 Run Way
Lee, Massachusetts 01238

Attn: Nick Bruzzi

Amanda Scott

Authorized for release by:
10/27/2022 12:53:34 PM

Amanda Scott, Project Manager
(574)233-4777
Amanda.Scott@et.eurofinsus.com

Table of Contents

Definitions/Glossary

Client: Housatonic Basin Sampling & Testing
Job ID: 810-41842-1
Project/Site: 1279000

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
¤	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000

Job ID: 810-41842-1

Job ID: 810-41842-1

Laboratory: Eurofins Eaton South Bend

Narrative

Job Narrative
810-41842-1

Receipt

The samples were received on 10/19/2022 9:00 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 0.8°C

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000

Job ID: 810-41842-1

Client Sample ID: POE Post GT Brook 01G

Lab Sample ID: 810-41842-1

PWSID Number: 1279000

Analyte	Result	Qualifier	RL	MDL	Unit	Dil Fac	D	Method	Prep Type
Perfluorooctanesulfonic acid (PFOS)	0.87	J	1.9	0.51	ng/L	1		537.1	Total/NA
Perfluorooctanoic acid (PFOA)	1.1	J	1.9	0.48	ng/L	1		537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	0.70	J	1.9	0.43	ng/L	1		537.1	Total/NA
Perfluorobutanesulfonic acid (PFBS)	0.89	J	1.9	0.69	ng/L	1		537.1	Total/NA

This Detection Summary does not include radiochemical test results.

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000

Job ID: 810-41842-1

Client Sample ID: POE Post GT Brook 01G
Date Collected: 10/11/22 09:50
Date Received: 10/19/22 09:00

Lab Sample ID: 810-41842-1
Matrix: Drinking Water
PWSID Number: 1279000

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	**0.87**	**J**	1.9	0.51	ng/L		10/21/22 07:02	10/22/22 12:42	1
Perfluoroundecanoic acid (PFUnA)	ND		1.9	0.61	ng/L		10/21/22 07:02	10/22/22 12:42	1
Perfluorohexanoic acid (PFHxA)	ND		1.9	0.61	ng/L		10/21/22 07:02	10/22/22 12:42	1
Perfluorododecanoic acid (PFDoA)	ND		1.9	0.61	ng/L		10/21/22 07:02	10/22/22 12:42	1
Perfluorooctanoic acid (PFOA)	**1.1**	**J**	1.9	0.48	ng/L		10/21/22 07:02	10/22/22 12:42	1
Perfluorodecanoic acid (PFDA)	ND		1.9	0.58	ng/L		10/21/22 07:02	10/22/22 12:42	1
Perfluorohexanesulfonic acid (PFHxS)	**0.70**	**J**	1.9	0.43	ng/L		10/21/22 07:02	10/22/22 12:42	1
Perfluorobutanesulfonic acid (PFBS)	**0.89**	**J**	1.9	0.69	ng/L		10/21/22 07:02	10/22/22 12:42	1
Perfluoroheptanoic acid (PFHpA)	ND		1.9	0.50	ng/L		10/21/22 07:02	10/22/22 12:42	1
Perfluorononanoic acid (PFNA)	ND		1.9	0.46	ng/L		10/21/22 07:02	10/22/22 12:42	1
Perfluorotetradecanoic acid (PFTeDA)	ND		1.9	0.63	ng/L		10/21/22 07:02	10/22/22 12:42	1
Perfluorotridecanoic acid (PFTrDA)	ND		1.9	0.58	ng/L		10/21/22 07:02	10/22/22 12:42	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		1.9	0.60	ng/L		10/21/22 07:02	10/22/22 12:42	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		1.9	0.49	ng/L		10/21/22 07:02	10/22/22 12:42	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		1.9	0.60	ng/L		10/21/22 07:02	10/22/22 12:42	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	ND		1.9	0.62	ng/L		10/21/22 07:02	10/22/22 12:42	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	ND		1.9	0.62	ng/L		10/21/22 07:02	10/22/22 12:42	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	ND		1.9	0.47	ng/L		10/21/22 07:02	10/22/22 12:42	1

Surrogate	*%Recovery*	*Qualifier*	*Limits*	*Prepared*	*Analyzed*	*Dil Fac*
13C2 PFHxA	*95*		*70 - 130*	*10/21/22 07:02*	*10/22/22 12:42*	*1*
13C2 PFDA	*92*		*70 - 130*	*10/21/22 07:02*	*10/22/22 12:42*	*1*
13C3 HFPO-DA	*90*		*70 - 130*	*10/21/22 07:02*	*10/22/22 12:42*	*1*
d5-NEtFOSAA	*87*		*70 - 130*	*10/21/22 07:02*	*10/22/22 12:42*	*1*

Method: EPA PFAS6 - PFAS6

Analyte	Result	Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
PFAS Total	ND		2.00	0.500	ng/L			10/27/22 14:30	1

Surrogate Summary

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000

Job ID: 810-41842-1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

		Percent Surrogate Recovery (Acceptance Limits)			
		PFHxA	PFDA	HFPODA	d5NEFOS
Lab Sample ID	Client Sample ID	(70-130)	(70-130)	(70-130)	(70-130)
810-41842-1	POE Post GT Brook 01G	95	92	90	87
LCS 810-36067/3-A	Lab Control Sample	97	97	96	88
LLCS 810-36067/2-A	Lab Control Sample	103	96	95	91
MBL 810-36067/1-A	Method Blank	108	99	99	86

Surrogate Legend

PFHxA = 13C2 PFHxA
PFDA = 13C2 PFDA
HFPODA = 13C3 HFPO-DA
d5NEFOS = d5-NEtFOSAA

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MBL 810-36067/1-A
Matrix: Drinking Water
Analysis Batch: 36138

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 36067

Analyte	MBL Result	MBL Qualifier	RL	MDL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanesulfonic acid (PFOS)	ND		2.0	0.53	ng/L		10/21/22 07:02	10/22/22 08:59	1
Perfluoroundecanoic acid (PFUnA)	ND		2.0	0.63	ng/L		10/21/22 07:02	10/22/22 08:59	1
Perfluorohexanoic acid (PFHxA)	ND		2.0	0.63	ng/L		10/21/22 07:02	10/22/22 08:59	1
Perfluorododecanoic acid (PFDoA)	ND		2.0	0.63	ng/L		10/21/22 07:02	10/22/22 08:59	1
Perfluorooctanoic acid (PFOA)	ND		2.0	0.50	ng/L		10/21/22 07:02	10/22/22 08:59	1
Perfluorodecanoic acid (PFDA)	ND		2.0	0.60	ng/L		10/21/22 07:02	10/22/22 08:59	1
Perfluorohexanesulfonic acid (PFHxS)	ND		2.0	0.44	ng/L		10/21/22 07:02	10/22/22 08:59	1
Perfluorobutanesulfonic acid (PFBS)	ND		2.0	0.71	ng/L		10/21/22 07:02	10/22/22 08:59	1
Perfluoroheptanoic acid (PFHpA)	ND		2.0	0.52	ng/L		10/21/22 07:02	10/22/22 08:59	1
Perfluorononanoic acid (PFNA)	ND		2.0	0.48	ng/L		10/21/22 07:02	10/22/22 08:59	1
Perfluorotetradecanoic acid (PFTeDA)	ND		2.0	0.65	ng/L		10/21/22 07:02	10/22/22 08:59	1
Perfluorotridecanoic acid (PFTrDA)	ND		2.0	0.60	ng/L		10/21/22 07:02	10/22/22 08:59	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		2.0	0.62	ng/L		10/21/22 07:02	10/22/22 08:59	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		2.0	0.51	ng/L		10/21/22 07:02	10/22/22 08:59	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		2.0	0.62	ng/L		10/21/22 07:02	10/22/22 08:59	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	ND		2.0	0.64	ng/L		10/21/22 07:02	10/22/22 08:59	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	ND		2.0	0.64	ng/L		10/21/22 07:02	10/22/22 08:59	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	ND		2.0	0.49	ng/L		10/21/22 07:02	10/22/22 08:59	1

Surrogate	*MBL %Recovery*	*MBL Qualifier*	*Limits*	*Prepared*	*Analyzed*	*Dil Fac*
13C2 PFHxA	*108*		*70 - 130*	*10/21/22 07:02*	*10/22/22 08:59*	*1*
13C2 PFDA	*99*		*70 - 130*	*10/21/22 07:02*	*10/22/22 08:59*	*1*
13C3 HFPO-DA	*99*		*70 - 130*	*10/21/22 07:02*	*10/22/22 08:59*	*1*
d5-NEtFOSAA	*86*		*70 - 130*	*10/21/22 07:02*	*10/22/22 08:59*	*1*

Lab Sample ID: LCS 810-36067/3-A
Matrix: Drinking Water
Analysis Batch: 36138

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 36067

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluorooctanesulfonic acid (PFOS)	191	180		ng/L		94	70 - 130
Perfluoroundecanoic acid (PFUnA)	191	171		ng/L		89	70 - 130
Perfluorohexanoic acid (PFHxA)	191	186		ng/L		97	70 - 130
Perfluorododecanoic acid (PFDoA)	191	164		ng/L		86	70 - 130
Perfluorooctanoic acid (PFOA)	191	186		ng/L		97	70 - 130
Perfluorodecanoic acid (PFDA)	191	168		ng/L		88	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	191	185		ng/L		97	70 - 130
Perfluorobutanesulfonic acid (PFBS)	191	187		ng/L		98	70 - 130
Perfluoroheptanoic acid (PFHpA)	191	184		ng/L		96	70 - 130
Perfluorononanoic acid (PFNA)	191	179		ng/L		94	70 - 130

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: LCS 810-36067/3-A
Matrix: Drinking Water
Analysis Batch: 36138

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 36067

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluorotetradecanoic acid (PFTeDA)	191	160		ng/L		84	70 - 130
Perfluorotridecanoic acid (PFTrDA)	191	167		ng/L		87	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	191	160		ng/L		83	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	191	160		ng/L		84	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	191	179		ng/L		94	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid	191	175		ng/L		92	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid	191	160		ng/L		84	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	191	178		ng/L		93	70 - 130

Surrogate	*LCS %Recovery*	*LCS Qualifier*	*Limits*
13C2 PFHxA	*97*		*70 - 130*
13C2 PFDA	*97*		*70 - 130*
13C3 HFPO-DA	*96*		*70 - 130*
d5-NEtFOSAA	*88*		*70 - 130*

Lab Sample ID: LLCS 810-36067/2-A
Matrix: Drinking Water
Analysis Batch: 36138

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 36067

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluorooctanesulfonic acid (PFOS)	1.92	1.87	J	ng/L		97	50 - 150
Perfluoroundecanoic acid (PFUnA)	1.92	1.67	J	ng/L		87	50 - 150
Perfluorohexanoic acid (PFHxA)	1.92	1.75	J	ng/L		91	50 - 150
Perfluorododecanoic acid (PFDoA)	1.92	1.50	J	ng/L		78	50 - 150
Perfluorooctanoic acid (PFOA)	1.92	1.61	J	ng/L		84	50 - 150
Perfluorodecanoic acid (PFDA)	1.92	1.75	J	ng/L		91	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	1.92	1.72	J	ng/L		90	50 - 150
Perfluorobutanesulfonic acid (PFBS)	1.92	1.71	J	ng/L		89	50 - 150
Perfluoroheptanoic acid (PFHpA)	1.92	1.60	J	ng/L		84	50 - 150
Perfluorononanoic acid (PFNA)	1.92	1.70	J	ng/L		89	50 - 150
Perfluorotetradecanoic acid (PFTeDA)	1.92	1.51	J	ng/L		79	50 - 150
Perfluorotridecanoic acid (PFTrDA)	1.92	1.64	J	ng/L		85	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	1.92	1.49	J	ng/L		78	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	1.92	1.63	J	ng/L		85	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	1.92	1.85	J	ng/L		96	50 - 150

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: LLCS 810-36067/2-A
Matrix: Drinking Water
Analysis Batch: 36138

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 36067

Analyte			Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits		
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid			1.92	1.60	J	ng/L		83	50 - 150		
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid			1.92	1.49	J	ng/L		78	50 - 150		
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)			1.92	1.74	J	ng/L		91	50 - 150		

Surrogate	*LLCS %Recovery*	*LLCS Qualifier*	*Limits*
13C2 PFHxA	*103*		*70 - 130*
13C2 PFDA	*96*		*70 - 130*
13C3 HFPO-DA	*95*		*70 - 130*
d5-NEtFOSAA	*91*		*70 - 130*

QC Association Summary

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000

Job ID: 810-41842-1

LCMS

Prep Batch: 36067

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-41842-1	POE Post GT Brook 01G	Total/NA	Drinking Water	537.1 DW	
MBL 810-36067/1-A	Method Blank	Total/NA	Drinking Water	537.1 DW	
LCS 810-36067/3-A	Lab Control Sample	Total/NA	Drinking Water	537.1 DW	
LLCS 810-36067/2-A	Lab Control Sample	Total/NA	Drinking Water	537.1 DW	

Analysis Batch: 36138

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-41842-1	POE Post GT Brook 01G	Total/NA	Drinking Water	537.1	36067
MBL 810-36067/1-A	Method Blank	Total/NA	Drinking Water	537.1	36067
LCS 810-36067/3-A	Lab Control Sample	Total/NA	Drinking Water	537.1	36067
LLCS 810-36067/2-A	Lab Control Sample	Total/NA	Drinking Water	537.1	36067

Analysis Batch: 36725

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
810-41842-1	POE Post GT Brook 01G	Total/NA	Drinking Water	PFAS6	

Lab Chronicle

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000

Job ID: 810-41842-1

Client Sample ID: POE Post GT Brook 01G

Lab Sample ID: 810-41842-1

Date Collected: 10/11/22 09:50

Matrix: Drinking Water

Date Received: 10/19/22 09:00

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	537.1 DW			36067	AB	EA SB	10/21/22 07:02
Total/NA	Analysis	537.1		1	36138	MH	EA SB	10/22/22 12:42
Total/NA	Analysis	PFAS6		1	36725	RD	EA SB	10/27/22 14:30

Laboratory References:

EA SB = Eurofins Eaton South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777

Accreditation/Certification Summary

Client: Housatonic Basin Sampling & Testing
Job ID: 810-41842-1
Project/Site: 1279000

Laboratory: Eurofins Eaton South Bend

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
Massachusetts	State	M-IN035	06-30-23

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
PFAS6		Drinking Water	PFAS Total

Method Summary

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000

Job ID: 810-41842-1

Method	Method Description	Protocol	Laboratory
537.1	Perfluorinated Alkyl Acids (LC/MS)	EPA	EA SB
PFAS6	PFAS6	EPA	EA SB
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA SB

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA SB = Eurofins Eaton South Bend, 110 S Hill Street, South Bend, IN 46617, TEL (574)233-4777

Sample Summary

Client: Housatonic Basin Sampling & Testing
Project/Site: 1279000

Job ID: 810-41842-1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
810-41842-1	POE Post GT Brook 01G	Drinking Water	10/11/22 09:50	10/19/22 09:00	1279000

SAMPLE COLLECTION RECORD | CHAIN OF CUSTODY | EURO

810-41842 Chain of Custody

PWS NAME:	SOUTHWICK WATER DEPARTMENT
PWS ID:	1279000
PWS TOWN:	Southwick
PWS CLASS:	COM

Housatonic Basin Sampling & Testing

80 RUN WAY
LEE, MA 01238
(413)248-4622

HBST P.O. # 1279000-221011

# of WO: ####

SAMPLE INFORMATION							FIELD RECORDED				MICRO BIOLOGY			CHEMICAL ANALYSIS												
ID	SAMPLE TYPE	BACTERIA DEP ID	Chem Sample ID	LOCATION DESCRIPTION	DATE/TIME	SAMPLER	Field Temp C°	Field pH	Field Turbitity (NTU)	Chl2 Res (Free)	BACTERIA HPC	BACTERIA 9223 P/A	BACTERIA 9223 QT	PFAS Blank												Preserved Na2S2o3
			[10000]	POE POST GT BROOK 01G -	10/11/22 9:50 AM	Tim Vreeland								X												

* also contains (2) 537 samples, along with Blank
10/19/22 mw

CUSTODY TRANSFER		DATE/TIME	NOTES
SAMPLER		10-18-22 1600	
RECEIVED		10/19/22 0900	
RELINQUISHED			
RECEIVED			
RELINQUISHED			

Note: Submit via EDEP unless designated Private or otherwise noted. Email report to: Admin@HousatonicBasin.com. Lab testing shall be in compliance with all State and Federal Drinking Water and applicable regulations.

Initial Temp: 0.4
Corrected Temp: 0.8
IR Gun # 29 wet

10/27/2022

Login Sample Receipt Checklist

Client: Housatonic Basin Sampling & Testing

Job Number: 810-41842-1

Login Number: 41842

List Source: Eurofins Eaton South Bend

List Number: 1

Creator: Wojcik, Mary

Question	Answer	Comment
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	